## Double pulses scheme for THz-XUV pump-probe experiment at FLASH

<u>Rui Pan</u>\*, Ekaterina Zapolnova, Torsten Golz, Karsten Klose, Siegfried Schreiber, and Nikola Stojanovic

Deutsches Elektronen-Synchrotron (DESY), Notkestrasse 85, Hamburg, D-22607, Germany

\*rui.pan@desy.de

At FLASH, the free electron laser in Hamburg, an unique scheme is used to laser X-ray pulse and THz pulse by the same electron bunch for pump-probe experiment. However the timing overlap is achieved at the price of losing pulse energy and bandwidth of the X-ray pulse in a back reflection, since the THz beamline is longer in optical path. A new scheme of lasing the X-ray and THz pulses are proposed here. To solve this issue, two electron bunches separated by 21.5 ns are generated. The first electron bunch is used to generate THz pulse and the second one is used to generate the X-ray pulse. The delay in the beamlines is compensated by the delay between the electron bunches. The feasibility of the double pulses scheme is demonstrated. In order to optimize the X-ray and THz pulses individually, suppressing the X-ray pulse energy down to one-tenth in the first bunch is achieved, and  $\sim 20$ fs (r.m.s.) timing jitter is measured by electro-optic spectral decoding method between the THz pulses generated by the two electron bunches.

## References

- Zapolnova, E., Golz, T., Pan, R., Klose K., Schreiber S. and Stojanovic, N., 2018. THz pulse doubler at FLASH: double pulses for pump-probe experiments at X-ray FELs, J. Synchrotron Rad. 25, 39–43
- [2] Gensch, M., Bittner, L., Chesnov, A., Delsim-Hashemi, H., Drescher, M., Faatz, B., Feldhaus, J., Fruehling, U., Geloni, G.A., Gerth, C. and Grimm, O., 2008. New infrared undulator beamline at FLASH. *Infrared Physics & Technology*, 51(5), pp.423-425.
- [3] Green, B., Kovalev, S., Asgekar, V., Geloni, G., Lehnert, U., Golz, T., Kuntzsch, M., Bauer, C., Hauser, J., Voigtlaender, J. and Wustmann, B., 2016. High-field high-repetition-rate sources for the coherent THz control of matter. *Scientific reports*, *6*.